
Reviewed: Jan 2025 Updated: Sept 2020





MRI WRIST WITHOUT CONTRAST

Scan Range	Variable according to area of interest/ body part.
FOV	Fits to patient/ body part.
Position	Wrist to be pronated and in superman position
Localizer	3 plane scout localizer



PLANE	SEQUENCE	SLT/ SP	FAT SAT	FOV	NOTES
Axial	T2	3 mm	Υ	8 - 10 cm	angle parallel to long axis of wrist from cor loc
Axial	T1	3 mm		8 - 10 cm	angle parallel to long axis of wrist from cor loc
Sagittal	T2	3 mm	Υ	10 cm	Perpendicular to axials
Coronal	T1	2 mm		10 cm	Angle parallel to radius / lunate/ capitate alignment
Coronal	T2	2 mm	Υ	10 cm	Angle parallel to radius / lunate/ capitate alignment
3D Coronal	FE			10 cm	Angle parallel to radius / lunate/ capitate alignment







Coronal Plane Axial Plane